

#### SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

#### SHAVIT SCALES LTD 7B Alenbi Street Haifa, Israel Asaf Yakuti Phone: 972 52 2266998

#### CALIBRATION

Valid To: August 31, 2024

Certificate Number: 6762.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1</sup>:

#### I. Mechanical

Parameter/Equipment	Range	CMC <sup>2, 4</sup> (±)	Comments
Balances <sup>3</sup>	Up to 20 000 g $R \le 0.5$ g	1 <i>R</i>	OIML/R 76, weights Class F1, M1 (OIML R111)
	Up to 200 kg $R \le 0.5$ kg	0.72 <i>R</i>	
	(200 to 500) kg $R \leq 0.5 \ \text{kg}$	0.83 <i>R</i>	
	(500 to 1000) kg (1000 to 1500) kg (1500 to 2000) kg (2000 to 3000) kg (3000 to 4000) kg (4000 to 5000) kg	0.99 <i>R</i> 1.3 <i>R</i> 1.5 <i>R</i> 1.7 <i>R</i> 2.2 <i>R</i> 2.4 <i>R</i>	

<sup>1</sup> This laboratory offers commercial calibration service and field calibration services.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, usually using a coverage factor of k = 2. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

(A2LA Cert. No. 6762.01) 07/08/2022

Page 1 of 2

<sup>3</sup> Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

<sup>4</sup> In the statement of CMC, *R* represents the resolution of the unit under test.

Page 2 of 2

(A2LA Cert. No. 6762.01) 07/08/2022





## **Accredited Laboratory**

A2LA has accredited

# SHAVIT SCALES LTD

Haifa, ISRAEL

for technical competence in the field of

### Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 8<sup>th</sup> day of July 2022.

Vice President, Accreditation Services For the Accreditation Council Certificate Number 6762.01 Valid to August 31, 2024

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.