SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

SHAVIT SCALES LTD
21 Hamaapilim St.
Kiryat Atta, Israel 2822878
Asaf Yakuti Phone: 972 72 2377777

CALIBRATION

Valid To: November 30, 2028 Certificate Number: 7381.01

In recognition of the successful completion of the A2LA evaluation process (including an assessment of the organization's compliance with A2LA's Calibration Program Requirements), accreditation is granted to this laboratory to perform the following calibrations[1, 5]:

I. Mechanical

Parameter/Equipment	Range	CMC[2, 4] (±)	Comments
Scales and Balances[3]	Up to 20 kg $R \leq 0.5$ g	1R	OIML/R 76, weights class F_1, M_1 (OIML R111)
	(>20 to 200) kg $R \leq 0.5$ kg	0.72R	
	(>200 to 500) kg $R \leq 0.5$ kg	0.83R	
	(500 to 1000) kg	0.99R	
	(>1000 to 1500) kg	1.25R	
	(>1500 to 2000) kg	1.5R	
	(>2000 to 3000) kg	1.66R	
	(>3000 to 4000) kg	2.2R	
	(>4000 to 5000) kg	2.4R	

[1] This laboratory offers commercial calibration service.

[2] Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

[3] Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

[4] In the statement of CMC, R represents the resolution of the unit under test.

[5] This scope meets A2LA's *P112 Flexible Scope Policy.*

Accredited Laboratory

A2LA has accredited

SHAVIT SCALES LTD

Kiryat Atta, ISRAEL

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated April 2017*).

Presented this 31st day of August 2026.

Mr. Trace McInturff, Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 7381.01
Valid to November 30, 2028

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.